

Universal Medical Engineers

Baapari Agan, City Polyclinic & diag. centre Building 2nd floor
Bohri Kadal Chowk Pin:- 190002, Srinagar.
e-mail:- universalmedicalengineers@gmail.com
Contact No:- 9622458106, 9797139262

INSTALLATION/SERVICE REPORT

REPORT No:- 525

Date:- 15 Oct 2022

Department & Location

Name Of Hospital

D.S. Healthcare (DS Healthcare)
Moomina Abad Anantnag J&K

State/Province:

Jammu and Kashmir

Contact Person:-

Mr Zaid

Tel:-

7006200036

Country:

INDIA

ITEM DETAILS

Model	Serial No.	Description
Autoclave		Installed machine checked Pressure
Horizontal Make	XXXXX	
Napstor Equipmt		OK, checked Autocut OK checked
11.235 Ltr		
9KW Load		Temp OK working in good condition
having a 3 drum		Handover to user in working
capacity full		
SS material		condition
Temp - 21°		
3 Phase		
		Warranty = 1 Year Comprehensive one Year
		consumable items not covers in warranty
		i.e any gasket and heating element
		Don't use Ground water

Installation Status

Complete

Incomplete


Further Action (IF Any)

Don't use ground water, keep voltage maintain 225 volt each Phase with Neutral and 0-5 volt earthing and damage due to negligence not covers in warranty

Engineer's Sign

Customer's Seal

[Signature]
15/10/2022



[Signature]

NOTE:

1. Installation has been carried out as per the items purchased and available on the installation site.
2. Power conditions observed and measured on the site are instantaneous
3. Customer is responsible for maintaining the required power and other operating conditions in accordance with the instruction manual.
4. We are not responsible if any fault will occur in the equipment due to High input voltage poor Grounding or power fluctuation especially during warranty period.